




PH-13 | 41329-010

Digital Mammographic Phantom NCCE



Four types of phantoms designed to evaluate CT-AEC performance
Image quality can be evaluated by noise and S.D. on the phantom section images



FEATURES

| Outer shape of the phantom simulates a compressed breast of D shape
| Targets includes simulated microcalcifications, nylon fibrils, acrylic disks, an aluminum ring, Teflon disks, a Teflon ruler (slope) and a resolution test chart

EVALUATION PARAMETERS

| Contrast resolution
| Frequency enhancement
| Noise and contrast transfer function

DESCRIPTIONS

SET INCLUDES


- 1 phantom
- 1 storage case

PH-10 | 41322-000

BMD Chart Phantom UHA



Bone Mineral Density chart for microdensitometry (MD) method



FEATURES

| 21 steps with different hydroxyapatite content
| Steps range from 0 to 400 mg/cm², with 20mg/cm² difference each

APPLICATIONS

| microdensitometry

DESCRIPTIONS

SET INCLUDES

- 1 chart phantom
- 1 storage case

